



Notes: *

- Sensor requires 10 gpm @ 15 psig of continuous flow.
- Control Box - 110/1/60 @ 0.4 Amps.
- Reuse Applications - a temporary source of liquid with high solids concentration is required for second calibration point. This would be liquid prior to sand filters, etc. with 4-10 ppm range. This is required at 10 gpm and 15 gpm at the sensor.

Options:

- Submersible pump, 1/2 hp, 110/1/60, 2" discharge for 10 gpm at 46" complete with fiberglass slide rail assembly for your 3/4" SS pipe slide rails. Intermediate slide rail supports required every 10'.
- Compressor - pancake compressor, 1 1/2 hp, 15 Amps, 110/1/60, 24" L x 24" W x 18" H, 7 CFM @ 60 psig.

Cerlic
 Cerlic Environmental Controls, Inc
 P.O. Box 420097
 Atlanta, GA 30342
 Phone (404) 256-3097
 Fax (404) 256-3094

FINAL EFFLUENT TSS METER
 ASSEMBLY DETAIL W/ SUN SHIELD
 P/N 24521B

NO.	REV.	DATE	BY	CHKD.	APP'D.
1					
2					
3					
4					
5					
6					
7					